

**Demonstration SET**  
**for**  
**METRIX DOX 2xxx**  
**Oscilloscopes**

**User's guide**

## Contents


HX0074 Test signal	Demo with			Page
	DOX2025	DOX2040	DOX2100	
n° 1 : Miscellaneous	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-2
n° 2 : Hystérésis	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-4
n° 3 : Pulse train	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-8
n° 4 : Data train-CS	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-10
n° 5 : Data frame - Fault	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-13
n° 6 : AM Modulation sinus	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-16
n° 7 : Square time base	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-18
n° 8 : Weak square with noise	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-19
n° 9 : Fast pulse comb	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-20
n° 10 : Digital frame-Fault	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-21
n° 11 : Frame - rare pulse	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-23
n° 12 : Recorder- 5 signals	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-24
n° 13 : Recorder heart	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-27
n° 14 : Harmonics	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-29
n° 15 : Distortion	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-30

GX1025 Signal	Demo with			Page
	DOX2025	DOX2040	DOX2100	
n° 1 : Long MEM with sinus 10MHz	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-31
n° 2 : Filters used on 2KHz square wave with a 62 kHz sine wave superposed	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-33
n° 3 : Filter and FFT sum of 2 sinus signals of 10 and 80kHz	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-36
n° 4 : Delayed time base on 2 sinus signals of 100 and 800 kHz	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-38
n° 5 : Delayed time base on 2 sinus signals of 10 and 80 kHz	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	II-41

# Demonstration of the functions of DOX2xxx Oscilloscopes using the HX0074 Demo Kit and the GX1025 Arbitrary Generator

## General description

- The HX0074 is an accessory with a circuit that generates 15 representative signals. It is associated with a guide describing the nature of the signals.
- The HX0074 demonstrator makes mastering the oscilloscope faster, because the display, analysis, and measurement of the signals generated by the HX0074 make use of all functions of the DOX2xxx.



- We will also use a GX1025 arbitrary generator to generate the signals specific to demonstrating the advantages of "LongMem" long recording memory depths and of using digital filters to observe composite signals.




## Presentation of the HX0074

- The HX0074 is built around a microprocessor. An LCD display unit and "UP/DOWN" buttons are used to select the desired signal. The HX0074 generates the signals on the "MAIN" and "AUX" BNCs. The HX0074 can be powered:
  - either by a standard 9V battery
  - or by a 12 VDC, 200mA external mains adapter, with positive polarity, that of METRIX MTX Mobile multimeters, for example. The power supply mode is selected using the switch.


<b>Demo:</b>	<b>with:</b>	<input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 1: Fanciful</b>				
<i>Nature</i>	4 pairs of successive signals approx. every 2 seconds.				
<i>Specs</i>	$2.6V < V_{pp} < 3.2V - 10Hz < F < 60Hz$				
<b>Oscilloscope Settings</b>	50 ms/div.-MAIN = CH1 = 1V/div.-AUX = CH2 = 1V/div.				
<i>Trigger</i>	standard on CH1 = MAIN				
<i>Modes</i>	XY (Display menu)				
<b>Objective</b>	Start in a playful manner by describing the different display modes: Normal, Delayed, XY				

a) Adjust the oscilloscope so as to display the signals correctly (possible using the "Autoset" key).




When Autoset is exited, the oscilloscope adjusts the vertical position so that the traces are not superposed.

b) Apply the "Delayed" and "ON" "Off" commands in succession to be able to observe a complete trace and zoom on a detail.




- c) Select the "XY mode" with CH1 on X and CH2 on Y.  
Observe that there is a succession of 4 geometrical shapes.




The geometrical shape obtained in XY depends on the sampling rate, which in our example is  $F_{\text{sample}} = 50 \text{ KHz}$

<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX 2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 2: Hysteresis</b>			
<i>Nature</i>	2 phase-shifted signals, triangle & pseudo-square			
<i>Specs</i>	Vpp $\approx$ 3.2V-F $\approx$ 1.7kHz- square-wave Tr $\approx$ 24 $\mu$ s-signal delay $\approx$ 40 $\mu$ s			
<b>Oscilloscope Settings</b>	100 $\mu$ s/div.-CH1 = MAIN = 500mV/div.-CH2 = AUX = 500mV/div.			
<i>Trigger</i>	standard on MAIN			
<i>Modes</i>	XY (Display menu)-no "Min/Max", and no "Repetitive Signal" (Horizontal menu)			
<b>Objectives</b>	"y(t)" and "XY" modes from phase-shifted signals Present automatic measurements with markers (F, square-wave Tr) Present phase measurements (manual, automatic) Present the FFT Mathematical function			

a) Adjust the oscilloscope so as to display the signals correctly (possible using the "Autoset" key) and select automatic phase measurement to determine the phase difference between the signals Main = CH1 and Aux = CH2.



b) Select the XY mode with CH1 on X and CH2 on Y.



The display of a hysteresis cycle is a "textbook case" often encountered in the educational context. It demonstrates the utility of display in the  $y(t)$  and XY mode(s), respectively.


Stress the simplicity of access to the XY mode, and of access to automatic phase measurement.

c) Return to "y(t) mode" in order to demonstrate the use of the automatic measurements (Ex: Vpp, Vamp, Freq, Rise,...).




d) Use of the FFT Mathematical function


The oscilloscope displays the CH1 signal and its FFT simultaneously. The "Time" cursors can be used to determine the frequencies of the fundamental and of the harmonics



The "Voltage" cursors can be used to determine the amplitude of the harmonics.



FFT of the signal on channel CH2



<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 3: Pulse train</b>			
<b>Nature</b>	1 signal containing trains of 10 pulses with variable spacing			
<b>Specs</b>	Vpp $\approx$ 3.4V-F $\approx$ 32kHz-Train spacings ( 100 to 180 $\mu$ s)			
<b>Oscilloscope Settings</b>	100 $\mu$ s/div.-CH1 = MAIN = 500mV/div.			
<b>Trigger</b>	on CH1 = MAIN-Hold-Off $\approx$ 354 $\mu$ s			
<b>Modes</b>	Triggered mode preferable			
<b>Objectives</b>	Triggering with "Hold-Off" on pulse trains			


a) Adjust the oscilloscope to display the signal on CH1 correctly (time base, sensitivity, and triggering source).

 Attention, with this type of signal, "Autoset" operation may be aleatory.

Without "Hold-Off", the triggering may act on any pulse of the train, as soon as the oscilloscope is ready to acquire.

This is accompanied by a feeling of "horizontal instability", making the display unusable.

The proper setting of the "Hold-Off" parameter (in the "Triggering" menu  $\rightarrow$  "Set") serves to ensure systematic triggering on the first pulse of the train.




This value must be greater than the duration of the pulse train, to disable triggering during this period, but must remain shorter than the time between 2 pulse trains (which varies between 400 and 480 $\mu$ s) .


In our case, the "Hold-Off" must be between 300 and 400 $\mu$ s.

b) Definition of the mask of the "Pass/Fail" function.

"Utility" → Pass/Fail → Mask Config → Mask Generation



c) Activation of the Pass/Fail function





The Pass/Fail function displays the number of times that the signal has satisfied ("Pass") or not satisfied ("Fail") the mask defined.

<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 4: Data+CS train</b>			
<b>Nature</b>	2 signals representing a digital frame (data) and a CS (chip select)			
<b>Specs</b>	Vpp $\approx$ 3.4V-F $\approx$ 40kHz (data)-F $\approx$ 1.5kHz (CS)			
<b>Oscilloscope Settings</b>	100 $\mu$ s/div.-MAIN = 1 V/div.-AUX = 1 V/div.			
<b>Trigger</b>	on BNC AUX = CH2			
<b>Modes</b>	Triggered mode preferable			
<b>Objectives</b>	Triggering on pulses			

a) Adjust the oscilloscope to display simply the 2 signals (time base, sensitivities and triggering source on the BNC AUX = CH2).


 Attention, with this type of signal, "Autoset" operation may be aleatory.

b) We are now going to demonstrate the utility of the pulse width triggers. The example chosen will serve to synchronize to the chip select signal of the data frame. We are going to trigger by turns on the width of the high level, then of the low level, of the "positive" pulse. In the first case, triggering will be on the negative-going edge of the chip select and in the second case it will be on the positive-going edge.




- c) Observe the first data group after the negative-going edge of the chip select using the "Delayed" function.  
 Example: to display the 2nd group of pulses, we shift the window by acting on the horizontal position

Expansion by 25



Expansion by 100



Horizontal displacement of the Zoomed zone by acting on the "Position" encoder:




<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 5: Data frame-Fault</b>			
<b>Nature</b>	2 signals representing a communication bus with "clock" & "data"			
<b>Specs</b>	Vpp ≈ 3.4V-F ≈ 31kHz (clock)-30μs < L+ < 200μs (data)			
<b>Oscilloscope Settings</b>	25μs/div.-MAIN = 1 V/div.-AUX = 1 V/div.			
<b>Trigger</b>	↙ on MAIN			
<b>Modes</b>	Triggered mode preferable-SPO mode, duration ≥2s			
<b>Objectives</b>	Capture and observe a rare event using SPO Triggering on pulse width of the AUX signal			

a) Adjust the oscilloscope so as to display the 2 signals in LongMem mode (time base, sensitivities, triggering source on MAIN).



*Attention, with this type of signal, "Autoset" operation may be aleatory.*

b) Observe a clock and the data bus using the "LongMem" function and the horizontal "Zoom".



The proposed signal is representative of a communication bus with "data - 8 bits" and a "clock". This communication setup is found in particular with the protocols of serial links such as the I2C bus, USB bus, CAN bus, Ethernet communication, etc.



The first utility of the operating mode is to detect and study faults in signals, without knowing their nature in advance, and therefore without having to set specific triggering conditions, for example.

In our example, we have frames approximately 3 ms apart and 1 frame in 120, or one frame every 360 ms, with the data at zero.


In STOP, using the horizontal zoom (x50) and choosing the position of the zoomed window, we can observe and analyze this frame and the one just before it and the one just after it.

Then, with the x1000 zoom factor, we observe the 6 clock pulses of the frame of zeros.

**Attention: the X1000 zoom factor is available only in the "LongMem" mode, which is available only on the DOX2040 and the DOX2100, on the DOX2025 or on the DOX2040-DOX2100 in Mem.depth mode = "Normal", the representation of the signal with the x1000 Zoom factor will be wrong.**




Zoom by 1000 in "LongMem" mode:




Zoom by 1000 in Memory Depth "**Normal**":

The representation is wrong: the train of 6 pulses is represented by a single pulse; the horizontal resolution is insufficient. This is because the sampling rate in Normal Memory Depth is 25KSPS, while it is 1MSPS in LongMem Memory Depth




<b>Demo:</b>	<b>with:</b>	<input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
<b>Test Signal</b>		<b>no. 6: Amplitude-modulated sine wave</b>			
	<i>Nature</i>	1 amplitude-modulated sinusoidal signal			
	<i>Specs</i>	1.3V < Vpp < 3.3V-F ≈ 1.3kHz			
<b>Oscilloscope Settings</b>		100μs/div.-MAIN = 500mV/div.			
	<i>Trigger</i>	on MAIN, 50% of Vpp			
	<i>Modes</i>	Triggered mode preferable-"Delayed" Mode			
<b>Objectives</b>		Display a fast-changing signal (e.g., modulation) Automatic "difference from reference" measurements			

Using the Delayed mode and the automatic peak amplitude measurement, we can observe the global shape of the signal and a zoomed zone. The measured amplitude (Vpp) of the signal in the zoomed zone is displayed on the right side of the screen. By shifting the zoom window using the "Horizontal position" button, we can determine the variation of the amplitude of the AM signal vs time.




Adjust the oscilloscope so as to display the signals correctly (possible using the "Autoset" function).


Persistence "Off"



Persistence Infinite




Difference from reference




We can press the "REF" key to record the signal on one of the 2 channels as reference, then validate this reference by "On" and observe the variations of the real-time signal of the channel with respect to the frozen reference.

<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 7: Square wave-Rise time</b>			
<i>Nature</i>	1 square wave, duty cycle 50%			
<i>Specs</i>	$V_{pp} \approx 3.4V$ - $F \approx 10kHz$ - $T_m \approx 800ns$			
<b>Oscilloscope Settings</b>	500ns to 200 $\mu s$ /div.-MAIN = 500mV/div.			
<i>Trigger</i>	 on MAIN, 50% of Vpp			
<i>Modes</i>	Triggered mode preferable-			
<b>Objectives</b>	Using automatic measurements (F, P, Tr, Tm, Vpp, Vrms, etc.) Activation of a particular measurement			

a) Adjust the oscilloscope so as to display the signal correctly (possible using the "Autoset" function), then validate the 23 automatic measurements available.



b) Rise Time measurement




<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 8: Square wave, low level, noisy</b>			
<b>Nature</b>	1 square wave of very low amplitude and very noisy			
<b>Specs</b>	5mV < Vpp < 30mV (depending on filtering)-F ≈ 1kHz			
<b>Oscilloscope Settings</b>	200 or 500µs/div.-MAIN = 2.5 or 5mV/div.			
<b>Trigger</b>	on MAIN, 50% of Vpp			
<b>Modes</b>	nothing at first, then 100kHz low-pass filtering			
<b>Objectives</b>	Triggering and display for a noisy signal Using the Digital filters			

a) Adjust the oscilloscope so as to display the signal approximately.




Attention, with this type of signal, "Autoset" operation may be aleatory.

The noisy signal is of low amplitude.



b) The use of a 100 kHz digital low-pass filter makes it possible to analyze the signal without the noise.




Demo:	with:	<input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 9: Comb of rapid pulses</b>				
<i>Nature</i>	Comb of 6 very brief pulses, with a low repetition rate				
<i>Specs</i>	$V_{pp} \approx 2V$ (depending on whether 50-Ohm load or not)- $F \approx 8kHz$				
<b>Oscilloscope Settings</b>	25 $\mu s/div.$ , then 10ns/div.-MAIN = 500mV/div.				
<i>Trigger</i>	 on MAIN, 50% of $V_{pp}$				
<i>Modes</i>	«LongMem», «Delayed», «ETS»				
<b>Objectives</b>	Utility of the ETS for an accurate and precise representation of signals "Delayed" & "LongMem" mode				

a) Adjust the oscilloscope so as to display the pulse trains and a zoomed train of 6 pulses.




Attention, with this type of signal, "Autoset" operation is in principle impossible.




Because of the very brief duration of the pulses (25ns) compared to their repetition interval ( $\approx 125\mu s$ ), we need a ratio of 1000 between the main time base and the "**Delayed**" time base.

b) We can also observe the train of 6 pulses in detail in the ETS mode with a 10ns/div time base. The example below presents a train of 6 pulses having a duration  $< 10ns$  with a rise time  $< 4ns$ :




Demo:	with:	<input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input type="checkbox"/> DOX2xxx
Test Signal	<b>no. 10: Digital frame+Fault</b>				
Nature	Digital frame with a recurrent fault				
Specs	F, square wave $\approx 5$ MHz, $V_{pp} \approx 1.8V-L+$ fault $\approx 7ns$				
Oscilloscope Settings	25 or 50ns/div., then 250ns/div-MAIN = 500mV/div., DC coupling				
Trigger	 DC coupling on MAIN, level $\approx 250mV$				
Modes	Select "Repetitive Signal" (Horiz menu)				
Objectives	Using triggering on pulse width				
	Using the LongMem and Delayed mode				


a) Display the signal (possible using the "Autoset" key), then set the parameters as indicated below. It can be seen that the display is not stable.



Set the triggering to pulse width < 20ns and adjust the triggering level to close to the low level of the pulse in order to trigger on the fault.




Use the Delayed and LongMem mode in order to be able to analyze the fault and the digital frame in detail:




<b>Demo:</b>	<b>with:</b>	<input checked="" type="checkbox"/> DOX2025	<input type="checkbox"/> DOX2040	<input type="checkbox"/> DOX2100	<input type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 11: Frame+Rare pulse</b>				
<b>Nature</b>	Digital clock signal with a fault				
<b>Specs</b>	F clock $\approx$ 5 MHz, Vpp $\approx$ 3.3V				
<b>Oscilloscope Settings</b>	100ns/div. then 25ns/div.-MAIN = 500mV/div., DC coupling				
<b>Trigger</b>	 DC coupling on MAIN, level $\approx$ 1.8V				
<b>Modes</b>	Triggered mode preferable-SPO mode, duration 1 or 2s				
<b>Objectives</b>	Capture and display of a rare fault in SPO mode Triggering possible on pulse width < 20ns, after SPO analysis				

- a) Adjust the oscilloscope so as to display the signal approximately (possible using the "Autoset" mode), then set the parameters as indicated opposite.
- b) The signal displayed represents a digital clock at 100ns.  
By paying attention, it may be possible to spot a certain instability of some edges of the signal.




- c) Enter Delayed mode with 10ns/div for the delayed time base.  
The fault is rather rare, since it affects only one clock pulse in 1000.  
It is a brief pulse, lasting less than 10ns, on the negative-going edge of the clock pulse.

We are going to use triggering on pulse width < 20ns by placing the triggering level on the top part of the clock pulse to achieve stable triggering on the fault, then use the Delayed & LongMem mode to display the clock signal and the fault.





<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 12: Recorder-5 signals</b>			
<i>Nature</i>	Tracking of 5 slow signals having varied shapes and characteristics			
<i>Specs</i>	Duration of each signal $\approx 1s$ , amplitude $1.5V < V_{pp} < 3.5V$			
<b>Oscilloscope Settings</b>	Duration-2s scale - $40\mu s$ -MAIN = $500mV/div.$ , DC coupling			
<i>Trigger</i>	None at first, then threshold(s) on MAIN, level according to signal			
<i>Modes</i>	"Source/Level" triggering,			
<b>Objectives</b>	Elementary presentation of the "Scan" mode for the $< 50ms/div$ time base Attention: the Scan and LongMem modes, along with the Scan and Delayed modes, are not compatible Attention: in Scan mode, for the "Level" to be active, the triggering must be in Normal mode when the Type of Trigger is on "Front"			


Observe in Scan mode the signals delivered by the HX0074 demonstrator:




Using triggering on pulse width and acting on the position of the level, it will be possible to trigger on each of the 5 slow signals.

By placing the triggering level close to zero and programming the pulse width, it is possible to synchronize to the damped sine wave, the low pulse, and the low ramp.






By placing the triggering level above the mean level of the signal and acting on the width of the positive pulse, it is possible to synchronize to the high pulse.




<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 13: Heart recorder</b>			
<i>Nature</i>	Slow "heartbeat" type signal & increasing/decreasing VDC			
<i>Specs</i>	Frequency of the signal $\approx 0.5s$ , amplitude $\approx 3.2V$ (heartbeat)			
<b>Oscilloscope Settings</b>	Duration 10s then 2s-MAIN & AUX = 500mV/div., DC coupling			
<i>Trigger</i>	None at first, then EXT thresholds on MAIN, levels 1V & 2.6V			
<i>Modes</i>	"Source/Level" triggering,			
<b>Objectives</b>	Entry of slow signals, "Normal" trigger mode			


a) Entry of signal no. 13 by "Front" type triggering, source CH2, "Normal" trigger mode.



b) Display of Signal no. 13 by Zooming in "STOP", serves to observe one period of signal CH2 and signal CH1 in detail.




c) The measurements can be made using the manual cursors, but it is also possible to display the 23 automatic measurements made on the desired channel simultaneously.





<b>Demo:</b>	with: <input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
<b>Test Signal</b>	<b>no. 14: Harmonics</b>			
<i>Nature</i>	2 signals, one square, the other triangular			
<i>Specs</i>	Frequency of the signal $\approx 50\text{Hz}$ , $V_{pp} \approx 3.2\text{V}$ (triangular), $V_{pp} \approx 3.4\text{V}$ (square)			
<b>Oscilloscope Settings</b>	5 ms/div.-MAIN & AUX = 1 V/div. DC coupling			
<i>Trigger</i>	↙ DC coupling on MAIN, 50% of $V_{pp}$ for example			
<i>Modes</i>	"Oscilloscope" mode, y(t) and			
<b>Objectives</b>	Display of a square and a triangular signal			

a) Adjust the oscilloscope so as to display the signal approximately in accordance with the first figure (possible using the "Autoset" mode), then set the parameters as indicated above. Then select the "FFT" Mathematical function.



Demo:	with:	<input checked="" type="checkbox"/> DOX2025	<input checked="" type="checkbox"/> DOX2040	<input checked="" type="checkbox"/> DOX2100	<input checked="" type="checkbox"/> DOX2xxx
Test Signal	<b>no. 15: Distortion</b>				
Nature	1 pseudo-sinusoidal signal containing harmonic distortion				
Specs	Frequency of the signal $\approx 50\text{Hz}$ , $V_{pp} \approx 3.2\text{V}$				
Oscilloscope Settings	2.5ms/div.-MAIN = 500mV DC coupling imperative				
Trigger	↙ DC coupling on MAIN, level 50% of $V_{pp}$ , for example				
Modes	"Oscilloscope" mode,				
Objectives	Display of a frequency-modulated 50Hz sine wave with components				

a) Adjust the oscilloscope so as to display the signal by pressing the "Autoset" key, then set the parameters as indicated above.



- ✓ Sine wave having an amplitude of 0.3 V (10%); frequency 150Hz (order 3); phase shift:  $\text{PI}$  ( $180^\circ$ )
- ✓ Sine wave having an amplitude of 0.6 V (18%); frequency 250Hz (order 5); phase shift:  $\text{PI}/2$  ( $90^\circ$ )


# 1. Using the SG1025 generator to demonstrate the advantages of the "LongMem" memory depth and of the Digital Filters

## a) Influence of memory depth (LongMem or Normal) on the sampling interval:


The sampling rate of the DOX2040-DOX2100 oscilloscopes for the time base position **M = 250 $\mu$ s/div**-for example- is **50MSPS** with a Memory Depth = «**LongMem**» and **2.5MSPS**. If the Memory Depth = "**Normal**", the minimum observable interval in DOTs will therefore be, in this case, 20ns when "LongMem" is active and 400ns in "Normal" mode.

To see the effect of the "**LongMem**" function during a fine analysis of a signal, we are going to observe, in "**Delayed**" mode, a **sinusoidal signal having a frequency of 10 MHz** in DOTs display mode and **Vectors** with a main time base of **M = 250 $\mu$ s/div** and a Delayed time base of **W = 25ns/div** (there is a ratio of **10000** between the **M** and **W** time bases):

When the LongMem function is activated, the 10 MHz sinusoidal signal is perfectly observable with the Delayed time base and a Zoom factor of 10000, because the 20ns sampling interval is shorter than the 50ns half-period of the sinusoidal signal



But when the memory depth is "Normal", the 10 MHz sinusoidal signal is no longer properly reconstituted, because the 400ns sampling interval is longer than the 100ns period of the 10 MHz sinusoidal signal:



Conclusion:


By switching the memory depth from "Normal" to "LongMem", we can record the same time interval with a recording interval 20 times as fine, making possible a finer analysis of the signal, in "Delayed" mode, for example.

## 2. Using the digital filters

### 2.1 2 kHz square wave with a 62 kHz sine wave superposed on its plateaus


a) Display of the 2 kHz square wave with a 62 kHz sine wave superposed on its plateaus:

Remark: the frequencies of the Digital Filters depend on the sampling frequency and therefore on the time base range ( $M = 250\mu\text{s}$ ), so we recommend observing the details of the signals with the "Delayed" time base ( $W = 25\mu\text{s}$ ) and a normal memory depth; this does not alter the cut-off frequencies of the filters, which depend on the main time base range,  $M = 250\mu\text{s}$ .




b) A "low-pass" digital filter having a high cut-off frequency of 25 kHz is applied to this signal:

The 62 kHz sinusoidal signal, which is above the high cut-off frequency of the filter, disappears and the edges of the 2 kHz square wave are rounded (see below).




c) A "high-pass" digital filter having a low cut-off frequency of 25 kHz is applied:

The 2 kHz square wave is blocked by the high-pass filter leaving only the 60 kHz sinusoidal signal, which is above the low cut-off frequency of the filter.




d) A "bandpass" digital filter having a pass band from 12.5 kHz to 100 kHz is applied:

The 2 kHz square wave, which is not in the pass band, is blocked by the filter, leaving only the 60 kHz sinusoidal signal, which is in the pass band



e) A "band-stop" digital filter (25 kHz to 800 kHz) is applied

The 62 kHz sinusoidal signal is attenuated, while the 2 kHz square wave passes intact.




## 2.2 Sum of 2 sinusoidal signals having frequencies of 10 kHz and 80 kHz


a) Display of the sum of 10 kHz & 80 kHz sine waves

The low frequency (~10 kHz) indicated by the hardware frequency counter is displayed at bottom left of the screen (take care to set the triggering level close to zero).

We use the cursors to measure the high frequency (~76.92 kHz)




When a "low-pass" digital filter having a cut-off frequency of 37.5 kHz is applied, the 10 kHz sine wave passes but the 80 kHz sine wave is highly attenuated:



We are now going to observe the effect of the "Low-Pass" filter on the 10 kHz/80kHz sine wave using the FFT.


Display of the Signal and its FFT with the "Off" digital filter:

The FFT shows the 10 kHz fundamental and the 80 kHz harmonic of the signal.



FFT of the signal with the low pass digital filter having a cut-off frequency of 37.5 kHz:

The FFT shows the 10 kHz fundamental but the 80 kHz harmonic has been highly attenuated by the digital filter



## 2.3 Product of 2 sine waves having frequencies of 100 kHz and 800 kHz

### Display of the product signal with Delayed time base

We use the cursors to measure the frequency of the product signal,  $F = 800$  kHz (Remark: the hardware frequency counter indicates 399.996 kHz because the triggering level is adjusted on the peaks of the product signal).


Configuration of the oscilloscope:

Time base range M = 100  $\mu$ s


Memory Depth = Normal

Acquisition: Samples


Display: Vectors



Then the low frequency of the product signal,  $F = 100$  kHz




The FFT of the product signal shows spikes at the sum frequency (900 kHz = 800 kHz + 100kHz) and at the difference frequency (700 kHz = 800 kHz - 100 kHz):




Using a bandpass filter centred on 700 kHz, then on 900 kHz, we are going to be able to isolate these 2 spectral components.


a) A bandpass digital filter (650 kHz 825 kHz) is used to isolate the 700 kHz spectral component:




Display of the 700 kHz spectral component with delayed time base:



b) A bandpass filter (825 kHz, 1 MHz) is used to isolate the 900 kHz component:



Display of the 900 kHz component with delayed time base:



## 2.4 Product of 2 sinusoidal signals having frequencies of 10 kHz and 80 kHz

### Display of the product signal with Delayed time base

We use the cursors to measure the frequency of the 80 kHz product signal (remark: the hardware frequency counter indicates 40.0 kHz because the triggering level is adjusted on the peaks of the product signal).


### Configuration of the oscilloscope:

Time base range M = 1.0ms


Memory Depth = Normal

Acquisition: Samples


Display: Vectors




Then the low frequency of the product signal, 10 kHz:






The FFT of the product signal shows two spectral components having the same amplitude, one at the  $F = 90\text{kHz}$  sum frequency ( $10\text{ kHz} + 80\text{ kHz}$ ) and the other at the  $F = 70\text{ kHz}$  difference frequency ( $80\text{ kHz} - 10\text{ kHz}$ ).




We use a bandpass digital filter (60 kHz 80 kHz) to isolate the 70 kHz spectral spike:





We view the filtered signal and use the cursors to measure its frequency,  $F = 70\text{ kHz}$ , and its amplitude =  $5.12\text{ V}_{\text{peak-to-peak}}$ :








We separate the 90 kHz spectral spike with a bandpass filter (85 kHz 100 kHz)





We view the filtered signal and use the cursors to measure its frequency,  $F = 90 \text{ kHz}$ , and its amplitude = 5.12 Vpeak-to-peak:





**Remark: the 70 kHz and 90 kHz components have the same amplitude.**