10µJ to 10J

Features

- Ø46mm apertures
- Metallic coating for high rep rates
- BF coating for highest damage threshold
- Rep rates up to 10kHz
- Measure lasers with pulse widths up to 20ms



PE50BF-C

Energy Sensor with optional heat sink

Jse perture mm bsorber Type Spectral Range µm ^(a) Surface Reflectivity % approx. Jalibration Uncertainty ±% ^(a) Max Pulse Width Setting ^(d) inergy Scales owest Measurable Energy µJ ^(c)	High rep rate Ø46 metallic 0.15 - 3 50 3					High dama	ae thresho	Id		
boorber Type Spectral Range µm ^(a) Surface Reflectivity % approx. Calibration Uncertainty ±% ^(a) Max Pulse Width Setting ^(d) Energy Scales	metallic 0.15 - 3 50					High damage threshold				
Spectral Range µm ^(a) Surface Reflectivity % approx. Calibration Uncertainty ±% ^(a) Max Pulse Width Setting ^(d) Energy Scales	0.15 - 3 50					Ø46				
Surface Reflectivity % approx. Calibration Uncertainty ±% ^(a) Max Pulse Width Setting ^(d) Energy Scales	50					BF				
Calibration Uncertainty ±% (a) Max Pulse Width Setting (d) Energy Scales						0.15 - 3, 10).6 ^(e)			
Iax Pulse Width Setting ^(d)	3					20				
nergy Scales	0					3				
	2µs	30µs	500µs	1ms	5ms	1ms	2ms	5ms	10ms	20ms
owest Measurable Eperavuul (c)	10J to 200µJ	10J to 200µJ	10J to 2mJ	10J to 2mJ	10J to 2mJ	10J to 2mJ	10J to 2mJ	10J to 20mJ	10J to 20mJ	10J to 20mJ
UWEST MEASULADIE ELIELYY HU	10	10	60	80	100	120	300	600	600	600
Nax Pulse Width ms	0.002	0.03	0.5	1	5	1	2	5	10	20
laximum Pulse Rate pps	10kHz	5kHz	900Hz	450Hz	100Hz	250Hz	100Hz	50Hz	40Hz	20Hz
loise on Lowest Range µJ	0.5	1	6	10	20	30	60	100	100	100
dditional Error with Frequency %	±2% to 2kHz ±4.5% to 5kHz	±2%	±2% to 750Hz	±2% to 400Hz	±1% to 80Hz	±1% to 100Hz ±2.5% to 150Hz ±4.5% to 250Hz	±1%	±1%	±1%	±2%
inearity with Energy for >7% of full scale	^{©)} ±1.5%					±2%				
Damage Threshold J/cm ² (b)										
<100ns	0.1					0.8				
1µs	0.2					1				
300µs	2					4				
2ms	6					10				
laximum Average Power W	15, 25 with optional heat sink					15, 25 with optional heat sink				
Maximum Average Power Density W/cm ²	20					20				
Iniformity over surface	±2% over central 50% of aperture					±2% over central 50% of aperture				
iber Adapters Available (see page 109)						ST, FC, SMA, SC				
Veight kg						0.25				
Compliance	CE, UKCA, China RoHS					CE, UKCA, China RoHS				
/ersion										
Part Number	7Z02936					7Z02934				
lote: (a) Calibration curve is verified and adjusted at pecified wavelengths. It other wavelengths, there may be an additional error p to the value given.	Specified wavelengths: 248-266nm, 355nm and 1064nm. Max additional error at 2940nm ±3%. Max additional error at other wavelengths: ±2%. <240nm not calibrated					Specified wavelengths: 193nm, 248-266nm, 355nm, 532nm and 1064nm. Max additional error at 2940nm ±3%. Max additional error at other wavelengths: ±2%.				

Note: (b)
For wavelengths below 600nm, derate damage threshold to 60% of given values. Below 300nm, derate to 40% of given values.

Note: (c) With the "user threshold" setting set to minimum. For other settings, the species for >7% of full scale or greater than twice the "user threshold", whichever is greater.

The user threshold is not available with LaserStar, Nova/Orion, Pulsar, USBI and Quasar. For these meters, the threshold is set to minimum and the linearity spec is >10% of full scale. The PE-C series will only operate with available with LaserStar, Nova/Orion, Pulsar, USBI and Quasar. For these meters, the threshold is set to minimum and the linearity spec is >10% of full scale. The PE-C series will only operate with available with LaserStar, Nova/Orion, Pulsar, USBI and Quasar. For these meters, the threshold is est to minimum and the linearity spec is >10% of full scale. The PE-C series will only operate with available with LaserStar, Nova/Orion with adapter can introduce up to 1% additional measurement error.

The user threshold feature allows adjustment of the internal threshold up to 25% of full scale if desired to avoid false triggering in noisy environments.

For further information, see the FAQs on our Website.


Note: (d) With the LaserStar, FUlsar, USBI, Quasar and Nova/Orion with adapter, only 2 out of 5 pulse widths settings are available; for the PE50-C model the 2µs (displayed as "10µs") and 1ms settings, and for the PE50BF-C model the 1ms and 10ms settings.


Note: (e) If the sensor is set to the 1064nm wavelength, then when measuring 10.6µm pulses, the reading will be approximately 1.19X the correct reading. If you use the attenuate function and set the attenuation to read 0.84, then you will have the correct reading at 10.6µm. The additional e

* For drawings please see page 106




1.2.3







PE25BF-DIF-C




PE50BF-DIFH-C / PE50U-DIFH-C



PE50-C / PE50BF-C



PE50BF-DIF-C / PE50-DIF-C



PE50BB-DIF-C

