SL 3101 Validation-Set for validation of large steam sterilizers

for large steam sterilizers



For the validation of steam sterilizers according to ISO 17665 as well as for the validation of washer-disinfectors and washer-disinfectors for endoscopes according to ISO 15883. The set can individually be expanded or compiled yourself from one or more data loggers (EBI 12 or EBI11), the appropriate interface and the software Winlog.validation.
Film: download

Product Description

The set contains:
- 5 x EBI 12-T471 Temperature data loggers with AL 107 silicone protection boxes
- 1 x EBI 12-TP453 Temperature/ pressure data logger with AL 101 silicone protection box
- EBI IF 200, 4-port Interface with USB connection and antenna
- Software Winlog.validation
- 2 x AL 190 silicone cable strap sets
- Carrying case „SYSTAINER“

Part No.: 1250-3101

Technical data

Options



AL 190
AL 190 Silicon cable strap

Part No. 1248-0190

Software

Winlog.pro Software

Winlog.pro - The general software for programming and evaluation of ebro data loggers



Part No. 1340-2355

Hardware requirements	Processor: Pentium 1 GHz, Main memory: 1 GB, Free hard disk space: 1 GB, USB
Software requirements	Windows Vista (32 bit or 64 bit), Windows 7,8,10 (32 bit or 64 bit)
Further requirements	Resolution: 1024x768

Winlog.med Software

Software (Standard-Software for the CSSD)



Part No. 1340-2363

Hardware requirements	Processor: Pentium 1 GHz, Main memory: 1 GB, Free hard disk space: 1 GB, USB
Software requirements	Windows Vista (32 bit or 64 bit), Windows 7,8,10 (32 bit or 64 bit)