


# VICAT APPARATUS

Model Number :GOTT-VA-01



## DESCRIPTION

GOTT-VA-01 consists of a metallic frame, graduated scale with index, sliding probe of 300 g, consistency plunger Ø 10 mm, glass base plate. (see accessories). 0 cc of mortar is comprised between 17 to 25 seconds.

## ADDITIONAL ITEMS

<b>HARDENED NEEDLE,DIA</b>	<b>CODE 300-359</b>
NEEDLE, hardened Ø 1 mm ASTM - AASHTO	
<b>FINAL NEEDLE,DIA</b>	<b>CODE 300-360</b>
FINAL NEEDLE Ø 1 mm (Standards: ASTM - AASHTO)	
<b>CONICAL PLASTIC MOULD DIA</b>	<b>CODE 300-361</b>
CONICAL PLASTIC MOULD Ø 60/70 h 40 mm (ASTM - AASHTO)	
<b>GLASS THERMOMETER-10 TO +50°C</b>	<b>CODE 300-362</b>

## Manuals :

- (1) All manuals are written in English
- (2) Model Answer
- (3) Teaching Manuals

## General Terms

- (1) Accessories will be provided where applicable.
- (2) Manuals & Training will be provided where applicable.
- (3) Designs & Specifications are subject to change without notice.
- (4) We reserve the right to discontinue the manufacturing of any product.

## Warranty :

2 Years

## ORDERING INFORMATION :

ITEM	MODEL NUMBER	CODE
VICAT APPARATUS	GOTT-VA-01	300-358

\* Proposed design only, subject to changes without any notice.