



Bluepard Instruments Co., Ltd.
Suzhou Being Medical Device Co., Ltd.

Vacuum Oven-LCD

Microprocessor controller (with timing function)

Vacuum degree Control:

Vacuum oven adopts the digital technique of simulated engineering to control vacuum degree.

The vacuum degree controlled by full-automatic electromagnetism valve which make the degree control more exact. Vacuum control rang: 20Pa ~9999Pa

Vacuum display rang: 1Pa ~9999Pa

Vacuum accuracy: 1Pa

Features:

- Programmable vacuum cycle
- LCD vacuum degree window
- Suitable for drying thermo-sensitive material
- Shorter drying time for sample that does not dry easily
- Dual layer tempered glass door for clear observation
- Fully automatic electromagnetism controlled vacuum
- Minimum heating time 50% less than traditional vacuum oven

