

Time Interval Meter (Micro Seconds)



ZMTIM μ S

Make	Zeal
Model	ZMTIM μ S
Input Power	230 VAC, 50 Hz / 110 VAC, 60 Hz
Test Conditions	For both Start & Stop Pulse by Decade Switches Test 1) Apply DC Voltage (Up to 24 VDC Max) Test 2) Remove Voltage Test 3) Closure of Contact Test 3) Open of Contact
Range	9999 Sec 999 Milli Sec 999 Micro Sec
Reset Facility	By Push Button
Display	16 X 2 Dot Matrix LCD Display
Resolution	1 Micro second
Accuracy	0.01 % $\pm$ 10 μ S
Protection	Start & Stop Pulse are isolated from each other
Weight	5 kg.