



Mask Air Leakproofness Tester GT-RA12-1



Application:

Air leakproofness test for full-face mask of self-priming filter type anti-particle respirator. Meet the test standard GB 2626-2006 standard for self-priming filtering anti-particle respirator for respiratory protective articles. Applicable to labor safety protection inspection centers, occupational safety inspection centers, disease prevention and control centers, respirator manufacturing enterprises, etc.

Standards:

GB 2626 Respiratory protective equipment self-priming filter type anti-particulate respirator
GB 2890 Respiratory protection self-suction filter gas mask

Key Specification

Model	GT-RA12-1
Air Source	Vacuum Pump
Test Head-form	1pc
Micro Pressure Gauge	Measure Range $\pm 2000\text{Pa}$
Timer	0~99.99s/min/h, can be set
Power Supply	220V 50Hz
Pumping Rate	2L/min